



# Municipal Corporation, Aurangabad.

## Health Department



Phone no. -0240-2333538-40  
Fax. No. 0240-2331213

No. AMC / Health /2019 / 17/19

Date :- 27.7.19

### दरपत्रक सूचना

प्रति,

विषय:- Dental chair with complete unit खरेदीकरिता दरपत्रक सादर करणे  
वरील विषयी आपणांस कळविण्यात येते की, Dental chair unit खरेदी करावयाचे आहे.  
आपण आपले दरपत्रक ८ दिवसात बंद पाकीटमध्ये सादर करावे.

Dental chair with complete unit खरेदी करावयाच्या तपशील खालील प्रमाणे आहे.

Sr.No.	Name of Equipment	Packing	quantity
1	Dental chair with complete unit with specifications	1	1

सदरील साहित्य सोबत दिलेल्या गुण निदेश गुणवत्तापूर्वक असावे आणि सोबत खालील दस्तावेज सादर करावे

अ.कं.	दस्तावेज
०१	पॅन काड सत्यप्रत
०२	व्हॅट नोंदणी
०३	शॉप अॅक्ट लायसन्स
०४	बँक डिटेल्स NEFT/ RTGS
०५	दरपत्रक रेट
०६	तीन वर्षातील ताळेबंद पत्रक

आपले दरपत्रक आरोग्य वैद्यकीय अधिकारी, औरंगाबाद महानगरपालिका औरंगाबाद यांच्या नावे सादर करावे.  
सदरील खरेदी किंवा दर पत्रक नाकारण्याचे अधिकार राखून ठेवण्यात आलेले आहेत.

आरोग्य वैद्यकीय अधिकारी,  
औरंगाबाद महानगरपालिका औरंगाबाद

Note:- Specifications attached.

शासकीय दंत महाविद्यालय व रुग्णालय, औरंगाबाद

सन 2018-19 या वित्तीय वर्षात राज्य योजनांतर्गत मंजूर अनुदानातून खरेदी करावयाच्या यंत्रसामुग्रीची विनिर्दिष्ट तक्ता  
विभागाचे नाव:- सामाजिकदंतशास्त्र विभाग

अ.क्र	यंत्रसामुग्रीचे नाव	आवश्यक नगसंख्या	यंत्रसामुग्रीची विनिर्दिष्टे
1	डेंटल चेअर विथ कम्प्लीट युनिट Dental Chair With Complete Unit	1	<p>A. Dental Chair (Swiss/German/Brazilian/France/Indian make) Electrically operated physiological dental chair having easy movement Up &amp;down movements &amp;back rest movements should be motorized Dental chair should have ergonomically stress free seat &amp;back rest. Dental chair should be adjustable with U bend grip to provide neck comfort Should have dynamic balance on a sturdy base with rust and corrosion resistant steel material.(Desirable net weight of the base 70 Kg or more) Dental chair should have Mounted with high &amp;low intensity of 9V/12V two intensity switch &amp;sensor operated HALOGEN LIGHT Dental chair should have Built in X ray viewer Dental chair should Be supplied with fixed stainless steel instrument tray Dental chair should have Ceramic spittoon of 22cm with appropriate water supply drainage system and conventional saliva ejector Dental chair should have High &amp;low vacuum motorized suction with flow control valve, auto drain &amp; auto flush system Two way switch facility for filling the tumbler &amp;flushing the basin bowl. It should have water filter single candle type for connection Dental chair should have Internationally accepted designs with material(steel &amp;other material) as per component of ISO.ISI accepted material Dental chair should Be supplied with euro molded stool having backrest, pneumatic piston for height adjustment. Dental chair should Be supplied with rubber floor mat Dental chair should have Two 3 way syringe for air &amp;water spray with autoclavable nozzles (one with control box) Modular system should be suspended</p>

Control box should be fitted with

1 airrotor point

2 3 way syringe

3 Micromotor speed 35000rpm brush type

Supplied with control box facility for smooth adjustable speed from 10,000-40,000 RPM with hand foot control variable speed as required.

d) Quick response with smooth rotation throughout, with facility of clockwise rotation should have both reverse and forward speed rotation

Compressor

1 HP and above

Noiseless and oil free

Tank capacity 20 liters and more

Pressure range 6-7-8 bars

Noise level 68 at 5 bars

Hand pieces (German/Brazilian/France/Swiss) It should be supplied with one air rotor handpiece / micromotor contra angle hand piece

A air rotor hand piece should be ultra push& super torque

B it should have small head ceramic bearing

C speed of air rotor up to 300000 rpm

D should be compatible with universal coupling

E contra angle hand piece with spare cartridges

E oil spray nozzle should be supplied with each hand piece

Stabilizer

1.5 KVA